

法定计量检定机构
Legal Metrological Verification Institute
无锡市计量测试院

Wuxi Institute of Metrology and Testing

校准证书

Calibration Certificate

证书编号: 19585205
Certificate No.

送检单位
Applicant

安科瑞电气股份有限公司

联络信息
Contact information

上海马陆镇马东工业园育绿路253号

计量器具名称
Name of Instrument

多用户系列电能表

型号 / 规格
Type/Specification

ADF400L AC3*220/380V AC3*1(6)A

出厂编号
Serial No.

72005250900002

制造单位
Manufacturer

安科瑞电气股份有限公司

接收日期
Date of Accept

2020年5月29日

授权签字人
Approved by

严海东

严海东

部长

核验员
Checked by

胡佳乐

胡佳乐

校准员
Calibrated by

何静涵

何静涵



校准日期 2020 年 5 月 29 日
Calibration Date Year Month Day

签发日期 2020 年 5 月 29 日
Issued Date Year Month Day



地址: 江苏省无锡市春新东路8号 电话(Tel): 0510-82803164 传真(Fax): 0510-82805098
Add: No.8 Chunxin East Road, Wuxi, Jiangsu
网址(HTTP): www.wxmtc.com 邮编(Post Code): 214101 电子邮件(Email): wxmtc@wxmtc.com





校准结果

Results of the calibration

有功基本误差 (%)							
输入值				输入电流负载 (%)			
U(V)	Ib(A)	相别	cosφ	100	50	10	5
220	6	ABC	1.0	+0.5	-0.10	-0.00	-0.00
			0.5L	+0.00	+0.10	+0.05	+0.00
			0.8C	-0.05	-0.00	+0.00	-0.05
220	6	A	1.0	-0.05	+0.10	+0.05	+0.00
			0.5L	+0.00	-0.05	-0.10	-0.05
220	6	B	1.0	-0.00	-0.00	+0.00	+0.05
			0.5L	+0.00	+0.10	+0.00	+0.10
220	6	C	1.0	-0.05	+0.05	+0.10	-0.00
			0.5L	+0.10	-0.00	-0.15	-0.05

所校准项目符合有功0.5S级技术依据要求

敬告:

Suggestions

1. 被校准仪器修理后, 应立即进行校准。

Recalibrate the instrument immediately after it has been repaired.

2. 在使用过程中, 如对被校准仪器的技术指标产生怀疑, 请重新校准。

Recalibrate the instrument when any suspicion about its performance arises.

3. 根据客户要求或校准文件的规定, 通常情况下12个月校准一次。

According to the client orders or the calibration documents, the calibration will be done every 12 months as a rule.

以下空白/Following blank